




CERTIFICATE OF CONFORMITY

Issuer's Name / Producer : GLASSCO LABORATORY EQUIPMENTS PVT. LTD.
Issuer's Address / Producer : Manglai, P.O. Khudda Kalan, Ambala Cantt - 133001
Haryana, India
Object of Declaration : Volumetric Flask,
Material : Borosilicate Glass
Purpose of Use : A type of laboratory flask, calibrated to contain a precise volume at a particular temperature. Volumetric flasks are used for precise dilutions and preparation of standard solutions.

Specification of flasks specified in Table-1, EN ISO 1042 : 1999

Catalogue No.	Product Description	N/S	Capacity mL.	Pack Qty	
130.202.01	Volumetric Flask, Batch Certified With PP Stopper	10/19	5	2	
130.202.02	Volumetric Flask, Batch Certified With PP Stopper	10/19	10	2	
130.202.02A	Volumetric Flask, Batch Certified With PP Stopper	10/19	20	2	
130.202.03	Volumetric Flask, Batch Certified With PP Stopper	10/19	25	2	
130.202.04	Volumetric Flask, Batch Certified With PP Stopper	12/21	50	2	
130.202.04B	Volumetric Flask, Batch Certified With PP Stopper	14/23	50	2	
130.202.05B	Volumetric Flask, Batch Certified With PP Stopper	12/21	100	2	
130.202.05	Volumetric Flask, Batch Certified With PP Stopper	14/23	100	2	
130.202.06	Volumetric Flask, Batch Certified With PP Stopper	14/23	200	2	
130.202.07	Volumetric Flask, Batch Certified With PP Stopper	14/23	250	2	
130.202.08	Volumetric Flask, Batch Certified With PP Stopper	19/26	500	2	
130.202.09	Volumetric Flask, Batch Certified With PP Stopper	24/29	1000	2	
130.202.10	Volumetric Flask, Batch Certified With PP Stopper	29/32	2000	1	
130.202.11	Volumetric Flask, Batch Certified With PP Stopper	34/35	5000	1	

The object of declaration described above is in conformity with the requirements of the following documents:

- DIN ISO 3585 Borosilicate Glass 3.3
- EN ISO 1042 Laboratory Glassware - Volumetric Flasks

Material: The above are made from DIN ISO 3585:1998, heat resistant Boro 3.3 Glass that meet ASTM E-438, TYPE-1, CLASS- A Standard.

Compliance: These Flasks comply in every respect characteristics, measures and accuracy are in full conformity With the provisions of the standard EN ISO 1042:1999.

Additional Information:

Glassco has installed the quality control system as per ISO 9001:2008 and products delivered to the market are in full conformity with the technical Specification as per standard and with all fundamental requirements to such products.

Certificate No. IPK 0039.69-01 issued by intertek certification limited, A UKAS Management System Accredited Body



This document has been produced electronically & is valid without signature